

PURCHASE OF LUXMETER

Ref: MDF/DL/03/01/24

1. THE ORDER SHOULD NOT BE PREFERABLY SPLITTED. THE COMPREHENSIVE QUOTE CONSIDERING AVAILABILITY OF ALL COMPONENTS, FOR AT CSIO WILL BE CONSIDERED.
2. THE REQUIRED QUANTITY IS MENTIONED AND COMPONENTS SHOULD BE SEALED/PACKED AS PER THE ORIGINAL FACTORY PACKING. IF FOUND DAMAGED, WILL BE RETURNED.
3. THE PRICE SHOULD BE INCLUDING EX-CSIO DELIVERY.
4. THE VALID COC SHOULD BE SHARED.
5. PAYMENT AND OTHER THINGS AS PER CSIO NORMS
6. PL MAIL AT harry.garg@csio.res.in, pme@csio.res.in
7. QUOTE TIME 3 DAYS FROM DATE OF TENDER.
8. THE DETAILS OF USER WILL NOT BE SHARED.
9. SUPPLIER/OEM TO SHARE THE DELIVERY TIMELINES AFTER PO AND PAYMENT TERMS.. PURCHASE OF COMPONENTS/ PO IS PURELY THE PREROGATIVE OF CSIO.

SPECIFICATIONS OF LUXMETER

1. Portable
2. Operating Humidity : < 100 % R.H
3. Range (Lux) : 0-50000
4. Power Supply: : 9 V Battery Approx. DC 2 mA. or equivalent
5. Display: 13 mm LCD Display or better
6. Operating Temperature: 0° to 50°C or better
7. Package Contains: Instruction Manual, Carrying Case
8. Dimensions (L x W X H) (mm): portable, handy
9. Sampling time: 0.4 Sec or better
10. Resolution: 1-100 Lux or better
11. Accuracy: $\pm 5\% + 2d$ or better
12. Weight (g): <250gm or less
13. calibration report (TBD)
14. Future calibration(TBD)

REQUIRED QUANTITY: 2NOS